



# FEBRILE POSITIVE CONTROL

## Proteinus control solutions for agglutination reactions with bacterial suspensions

### REAGENTS

Salmonella Positive control:	Non-titered polyvalent antiserum produced in rabbits suited to be used as positive control in agglutination reactions with bacterial suspensions Salmonella in tube, in microplate and in slide.
Brucella Positive control:	Non-titered polyvalent antiserum produced in rabbits suited to be used as positive control in agglutination reactions with bacterial suspensions Brucella in tube, in microplate and in slide.
Proteus Positive control:	Non-titered polyvalent antiserum produced in rabbits suited to be used as positive control in agglutination reactions with bacterial suspensions Proteus in tube, in microplate and in slide.

### REAGENTS PREPARATION

Reagents are ready to use.  
Stability: until expiration date on label, stored at 2-8°C.  
Do not freeze.

### PRECAUTION

Reagent may contain some non-reactive and preservative components.  
It is suggested to handle carefully it, avoiding contact with skin and swallow.  
Perform the test according to the general "Good Laboratory Practice" (GPL) guidelines.

### NOTE

Only for IVD use.

### WASTE DISPOSAL

Product is intended for professional laboratories.  
Waste products must be handled as per relevant security cards and local regulations.

### PACKAGING

#### CODE BS00011

Salmonella positive control	1 x 0.5 ml
Brucella positive control	1 x 0.5 ml
Proteus positive control	1 x 0.5 ml

### REFERENCES

Edward J Young. Clinical Infectious Diseases 1995; 21: 283-290.  
David A et al. Current Opinion in Infectious Diseases 1994; 7: 616-623.  
David R et al Current Opinion in Infectious Diseases 1993; 6: 54-62.  
Chart H. Serodiag Immunother Infect Diseases 1995; 7: 34-39.  
Chart H. Serodiag Immunother Infect Diseases 1995; 7: 30-33.  
Bradley D Jones. Annu Rev Immunol 1996; 14: 533-561.

### MANUFACTURER

LTA s.r.l.  
Via Milano 15/F 20060  
Bussero (Milan) ITALY  
Tel: ++39 02 95409034  
Fax: ++39 02 95334185  
e-mail: [info@ltaonline.it](mailto:info@ltaonline.it)  
Website: <http://www.ltaonline.it>

### SYMBOLS

	Only for IVD use
	Lot of manufacturing
	Code number
	Storage temperature interval
	Expiration date (year, month)
	Warning, read enclosed documents
	Read the directions
	Biological risk

Mod. 01.06 (ver 2.2 - 27/09/2007)

